



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
 LR1355

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Ion Science Ltd. (Manufacturer)
Address	The Way, Flowmere Cambridge UK SG8 7UJ
Applicant	Elok-Opava, spol. s r.o.
Address	Sadek 17, 747 75 Velké Heraltice Czech Republic
Project Number	LR1355-1
Product	Benzene Monitor
Model Number	EWF1.1
Ratings	Ex d IIB+H2 T4 Gb AEx d IIB+H2 T4 Gb Ta= -20°C to +55°C 19-32VDC, 4-2.4A, IP65
Applicable Standards	CAN/CSA-C22.2 No. 61010-1-12 3rd ed CAN/CSA-C22.2 No. 60079-0:15 CAN/CSA-C22.2 No. 60079-1:11 UL 61010-1 3rd ed UL 60079-0 6th ed UL 60079-1 7th ed
Factory Location	Same as Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams, P.Eng.
 Manager, Hazardous Locations Dept. [Ex. Equipment]

Signature: 

Date: April 22, 2016