



防爆合格证

CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

证号
Certificate No. CE13. 2231X

产品名称
Name of Product
型号及规格
Type of Product
防爆标志
Marking
技术文件
Technical Documents
图号
Drawing No.
备注
Note (s)

红外线测温仪

FLUKE 568 EX

Ex ia II C T4 Gb

/

/

1. 产品采用两节 Varta Max Tech / Varta Industrial / Duracell Plus / Panasonic Pro Power AAA 电池串联供电。
2. 产品本身不允许带入 0 区, 但允许对 0 区执行非接触测量。
3. 产品使用时手柄必须带有 LH568 型皮质外套。
4. 环境温度: $0^{\circ}\text{C} < T_a < 50^{\circ}\text{C}$ 。

经对上述产品图样及技术文件的审查和样品的检验, 其符合以下标准:

By verifying the drawings and technical documents and checking samples, the product complies with the following standards:

GB3836.1-2010

GB3836.4-2010

发给: ecom instruments GmbH
Issued to:

本证失效日期: 2018-11-18
Date of Expire:

发证日期: 2013-11-18
Date of Issue:

中心印章
Center seal



中心主任
Director

刘红光

石油和化学工业电气产品防爆质量监督检验中心
Supervision & Test Center of Ex-products of China Petroleum & Chemical Industry

注: 本证仅对与送检样品一致的产品有效。

Note: This certificate is only valid for the products that are in accord with sample(s) tested and verified.

中心地址: 中国天津市丁字沽三号路 85 号

Center Add: No.85 No.3 Road DingZiGu Tianjin China Post code: 300131

邮政编码: 300131

电话/传真: 022-26651066 / 26689116

Tel/ Fax: 022-26651066/26689116

http://www.pcec.com.cn



CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

Certificate No.: CE13.2231X

Name of Product: Infrared Thermometer

Type of Product: FLUKE 568 EX

Marking: Ex ia II C T4 Gb

Technical Documents: /

Drawing No.: /

Note (s):

1. The product is powered by two Varta Max Tech / Varta Industrial / Duracell Plus / Panasonic Pro Power AAA batteries in series.
2. The product is not allowed to bring the device itself into zone 0 but allowed to measure contactless into zone 0.
3. The product shall only be used with the leather wrapper (LH568) around the hand grip.
4. The ambient temperature is $0^{\circ}\text{C} < T_a < 50^{\circ}\text{C}$.

By verifying the drawings and technical documents and checking samples, the product complies with the following standards:

GB3836.1-2010

GB3836.4-2010

Issued to: ecom instruments GmbH

Date of Expire: 2018-11-18

Date of Issue: 2013-11-18

Center seal



Director


Liu Hongguang

Supervision & Test Center of Ex- products of China Petroleum & Chemical Industry

Note: This certificate is only valid for the products that are in accord with sample(s) tested and verified.

Center Add: No.85 No.3 Road DingZiGu Tianjin China Post code: 300131
E-mail: cec@pcec.com.cn

Tel/ Fax: 022-26651066/26689116
<http://www.pcec.com.cn>